



oomar, INC.
MANUFACTURERS OF HIGH PRECISION OPTICS

GLEN COVE, L. I., NEW YORK 11542

Telephone: ORiole 6-1900

August 22, 1967

Miss Laura Gilpin
409 Camino del Monte Sol
Santa Fe, New Mexico 87501

Reference: S.E.I. Meter for Repair

Dear Miss Gilpin:

Thank you for sending in your S.E.I. Meter, Serial #339168A, for repair. After checking this instrument we find the following:

Must replace Photocell, Contact Plate,
Rheostat wiper, and two small lower scales.
Clean, lubricate and recalibrate meter.

The cost of this repair including parts and labor is \$ 36.50, F.O.B. Glen Cove.

We will await your authorization before proceeding with the work required.

Sincerely,

ZOOMAR, INC.

Walter Steuer /cac

Walter Steuer
Vice President

WS:cac